
Carbon Monoxide co



Performance

Measuring range	5 to 50 ppm		
Number of pump strokes	2 (200 ml)		
Correction factor	1		
Sampling time	6 min		

 $\begin{array}{lll} \mbox{Detecting limit}: & \mbox{1 ppm } (2 \mbox{ pump strokes}) \\ \mbox{Colour change}: & \mbox{Yellow} \rightarrow \mbox{Blackish brown} \end{array}$

Corrections for temperature & humidity: Unnecessary

Relative standard deviation: 15 % (for 5 to 10 ppm), 10 % (for 10 to 50 ppm)

Shelf life: 3 years

Reaction principle

 $CO + Na_2Pd(SO_3)_2 \rightarrow Pd + CO_2 + SO_2 + Na_2SO_3$

Possible coexisting substances and their interferences

Substance	Concentration	Interference	Changes colour by itself to
Acetylene	≥ 1/50	+	<u> </u>
Hydrogen sulphide	≥ 1/5	+	Blackish brown
Sulphur dioxide	≥ 1/10	+	J
Nitrogen dioxide	≥ 40 ppm	+ (Bleaching)	Red

Calibration gas generation

High pressure gas cylinder method

Special note

The demarcation of colour change layer might not be clear. If this is the case, read the tube at the demarcation (NOT at the middle point between the dark layer end and the pale layer end).