

**CERTIFICATE OF ANALYSIS**

PRODUCT BR0052G
RINGERS SOLUTION TABLETS (100S)

LOT NUMBER 4302835

EXPIRY DATE 2024.02.01

DATE OF MANUFACTURE 2021.02.01

Delivery/Customer information

Date Printed
2022.06.08

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	White tablet	White tablet
Colour on reconstitution	Colourless	Colourless
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery %	Test Description
-----------------------------	-------------	----------	------------	------------------

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at $37 \pm 2^\circ\text{C}$ for 18 ± 2 hours

Tested as a diluent

A satisfactory result is represented by recovery of $\pm 30\%$ of the control cfu (0 minutes) after holding at $20\text{-}25^\circ\text{C}$ for 45 minutes to 1 hour for *Escherichia coli* and *Staphylococcus aureus*.

<i>E. coli</i> ATCC@25922 WDCM00013	69	56	81	White/grey cols
<i>E. coli</i> ATCC@8739 WDCM00012	72	61	85	White/grey cols
<i>S. aureus</i> ATCC@25923 WDCM00034	150	150	100	White/grey cols

Control Medium: Tryptone Soya Agar

Refer to product specification for full details.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT BR0052G
RINGERS SOLUTION TABLETS (100S)
LOT NUMBER 4302835

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.
Lot Accepted. 2022.04.25

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com